

JGL d.d.
Svilno 20
51000 Rijeka
OIB: 20950636972

Rijeka, 01. lipnja 2023.

SVIM ZAINTERESIRANIM GOSPODARSKIM SUBJEKTIMA

Predmet: Odgovor na upit u postupku nabave s obveznom objavom za predmet nabave: Nabava analitičkog instrumenta za kontrolu proizvoda tokom proizvodnog procesa, Evidencijski broj nabave: EV-34

Naručitelj je dana 31. svibnja 2023. godine zaprimio upit zainteresiranog gospodarskog subjekta u predmetnom postupku nabave sljedećeg sadržaja:

1. UPIT:

Vezano na vaš projekt koji imate otvoreni za:

Naziv projekta: RINO SPREJ

Broj ugovora: KK.01.2.1.02.0111, Datum objave 26.05.2023

Javljam se s pitanjem vezano pod Tehničkim zahtjevima stavka 6.2.

Točka 6.2 odnosi se na procesni priključak gdje je definirano DN25. Prema vašim specifikacijama traži se procesni priključak DN25, ali prema dosadašnjim iskustvima s instalacijama obično se koristio tri-clamp 2".

Molim za informaciju koji procesni priključak je potrebno ponuditi.

1.ODGOVOR NARUČITELJA:

Procesni priključak može biti DN25 prema tehničkoj specifikaciji, ali prihvaćamo i tri-clamp 2". Priključak ovisi o mjestu na kojem se sonda instalira na mješač te imamo obje opcije priključaka.



Europska unija
Zajedno do fondova EU

Prethodni odgovor i I. izmjenu poziva na dostavu ponude s novim rokom za dostavu ponuda, Naručitelj javno objavljuje na istim internetskim stranicama na kojima je objavljen i Poziv na dostavu ponude te isti čini sastavni dio natječajne dokumentacije.

Naručitelj produljuje rok za dostavu ponuda do 09. lipnja 2023. godine do 10:00 sati.

Naručitelj poziva sve zainteresirane gospodarske subjekte da pročitaju i prate sva pojašnjenja koja su objavljena u sklopu ovog postupka nabave.

Odbor za nabavu

*JGL d.d.
Svilno 20
51000 Rijeka
Registration Number: 20950636972*

Rijeka, June 1st 2023.

TO ALL INTERESTED BUSINESS ENTITIES

***Subject:** Response to the request in the procurement procedure with mandatory publication for the subject of procurement: Procurement of an analytical instrument for product control during the production process, procurement number: EV-34*

On May 31st 2023, the contracting authority received a request from an interested economic entity in the subject procurement procedure with the following content:

1. REQUEST:

*Related to your project that you have open for:
Project name: RINO SPRAY
Contract number: KK.01.2.1.02.0111, Date of publication 05/26/2023*

I am contacting you with a question related to the Technical requirements of paragraph 6.2. Point 6.2 refers to the process connection where DN25 is defined. According to your specifications, a DN25 process connection is required, but according to previous experience with installations, a tri-clamp 2" was usually used.

I am asking for information on which process connection should be offered.

THE CONTRACTING AUTHORITY'S RESPONSE:

The process connection can be DN25 according to technical specification, but we also accept tri-clamp 2". The connection depends on the location where the probe is installed on the mixer and we have both port options.



Europska unija
Zajedno do fondova EU

The Contracting Authority publishes the previous response on the same website where the Tender Submission Invitation was published and it is an integral part of the tender documentation.

The Contracting Authority extends the deadline for tenders submission until June 9th 2023 until 10:00.

The Contracting Authority invites all interested economic operators to read and follow all the explanations published as part of this procurement procedure.

Procurement Committee